



EU - TYPE EXAMINATION CERTIFICATE

No. SK 19 - 088 MI-003 Rev. 4



This revision replaces all previous versions of this Certificate in full wording

Issued by	Slovenská legálna metrologia, n. o. Hviezdoslavova 31 974 01 Banská Bystrica Slovak Republic	Notified Body number 1432
In accordance with	Annex II, Module B to Regulation of the Government of the Slovak Republic No 145/2016 Coll. on making available of measuring Instruments on the market, in the wording of the Regulation of the Government of the Slovak Republic No 328/2019 Coll., which implements, in Slovakia, the Directive 2014/32/EU of the European Parliament and of the Council of 26 February 2014 on the harmonisation of the laws of the Member States relating to the making available on the market of measuring instruments as later amended (MID).	
Applicable essential requirements	Annex I and Annex V to MID	
Manufacturer	ZPA Smart Energy a.s. Komenského 821 541 01 Trutnov, Česká republika	
Applicant	Manufacturer	
Measuring instrument	Active electrical energy meter	
	Type	AM37x...
	Trade mark	see Descriptive annex
	SW version	see Descriptive annex
	Environment classes	
	- climatic	(-40 to +70) °C
	- mechanical	M1
	- electromagnetic	E2
Description and documentation	The principal technical and metrological data, characteristics, instrument description and approval conditions are set out in the Descriptive annex to this EU - type examination certificate (14 pages), which is part of this EU - type examination certificate. The test reports, designs, schematic diagrams and documentation used during certification process are stored in the reference folder ZPA_AM37x_00 through 04.	

Valid until **21 February 2029**

Date of issue **24 August 2022**



Ing. Štefan Král, PhD.
Representative of Notified Body



Where the instrument is subject to other Directives covering other aspects, this EU - type examination certificate is valid, assuming that the instrument conforms to the provisions of those Directives. Without written permission of the notified body this certificate may be reproduced only as a whole.